

Scan bodies, model analogs and printed models – Compatibility chart

Atlantis FLO, Atlantis IO FLO and Elos Accurate IO Scan Body can be used with all connectivity when scanning for Atlantis abutments. All products in the chart are included in the Atlantis Implant Kit for 3Shape (DME file). This chart is valid for DME version 1 or higher (*Dentsply Sirona Implants Atlantis abutments_1_20XX.dme*).

Note! 3Shape Scan Bodies can only be used when sending in the scan files using the validated direct workflow from 3Shape to Atlantis WebOrder. The scan body

detection must be done in the 3Shape software, therefore it is NOT possible to scan with 3Shape Scan Bodies, export the .stl files and send in the files utilizing Atlantis Scan Upload or any other .stl upload.

Atlantis Printed Model is available to order in Atlantis WebOrder and includes the Elos Accurate Analog for Printed Models. For MIS implants the MIS analog is included.



Manufacturer	Implant	Model scanning		Intraoral scanning						
		Atlantis FLO		Atlantis IO FLO	Elos Accurate IO Scan Body	3Shape Scan Body <small>Only Atlantis Implant Kit for 3Shape (DME)</small>	Analog for Printed Models		Atlantis Printed Model	
		Code (REF)	Kit Select	Code (REF)	REF	REF	Elos REF	MIS REF	Supported	
Dentsply Sirona	EV Connection¹									
	EV (XS) Green	P-01 (34593)	x	IO-P-01 (35243)	IO 3B-A	DEG	PMA-ATE30-1 ³	—	YES	
	EV (S) Purple	P-02 (34594)	x	IO-P-02 (35244)	IO 3B-B	DEP	PMA-ATE36-1 ³	—	YES	
	EV (M) Yellow	P-03 (34595)	x	IO-P-03 (35245)	IO 3B-C	DEY	PMA-ATE42-1 ³	—	YES	
	EV (L) Blue	P-04 (34596)	x	IO-P-04 (35246)	IO 3B-D	DEB	PMA-ATE48-1 ³	—	YES	
	EV (XL) Brown	P-05 (34597)	x	IO-P-05 (35247)	IO 3B-E	DE5	PMA-ATE54-1 ³	—	YES	
	Astra Tech Implant EV Profile									
	4.2 Yellow	P-06 (34598)	x	IO-P-06 (35248)	IO 3C-A	—	PMA-ATE42P-1	—	YES	
	4.8 Blue	P-07 (34599)	x	IO-P-07 (35249)	IO 3C-B	—	PMA-ATE48P-1	—	YES	
	Astra Tech Implant System TX									
	OsseoSpeed / OsseoSpeed TX 3.0 – Yellow	C-01 (34530)	x	IO-C-01 (35215)	IO 3A-A	DAX	PMA-ATO30-1	—	YES	
	OsseoSpeed / OsseoSpeed TX 3.5, 4.0 – Aqua	C-02 (34531)	x	IO-C-02 (35216)	IO 3A-B	DAS	PMA-ATO35-1	—	YES	
	OsseoSpeed / OsseoSpeed TX 4.5, 5.0 – Lilac	C-03 (34532)	x	IO-C-03 (35217)	IO 3A-C	DAL	PMA-ATO45-1	—	YES	
	OsseoSpeed TX Profile 4.5, 5.0	C-04 (34592)	x	IO-C-04 (35218)	—	—	—	—	—	

Please turn over →

All trademarks and company names are the property of their respective owners.

² The Atlantis IO FLO for Ankylos C/X is for single use only.

³ Use together with Elos lab screw.

⁴ Scanbody/Model analog exist but is not supported for Atlantis solutions.

Manufacturer	Implant	Model scanning		Intraoral scanning					
		Atlantis FLO		Atlantis IO FLO	Elos Accurate IO Scan Body	3Shape Scan Body Only Atlantis Implant Kit for 3Shape (DME)	Analog for Printed Models		Atlantis Printed Model
		Code (REF)	Kit Select	Code (REF)	REF	REF	Elos REF	MIS REF	Supported
Dentsply Sirona	Xive / Frialit								
	Xive S 3.0	J-01 (34575)	x	IO-J-01 (35230)	IO 8A-A	DX3	PMA-DXI30-1	—	YES
	Xive S 3.4	J-02 (34576)	x	IO-J-02 (35231)	IO 8A-B	DXG	PMA-DXI34-1	—	YES
	Xive S 3.8	J-03 (34577)	x	IO-J-03 (35232)	IO 8A-C	DXY	PMA-DXI38-1	—	YES
	Xive S 4.5	J-03 (34577)	x	IO-J-04 (35264)	IO 8A-D	DXB	PMA-DXI45-1	—	YES
	Xive S 5.5	J-03 (34577)	x	IO-J-05 (35265)	IO 8A-D	DXR	PMA-DXI55-1	—	YES
	Frialit 3.4	J-02 (34576)	x	—	—	—	—	—	—
	Frialit 3.8, 4.5, 5.5	J-03 (34577)	x	—	—	—	—	—	—
Ankylos									
	Ankylos C/X 3.5, 4.5, 5.5, 7.0	L-01 (34583)	x	IO-L-01 (35235) ²	IO 8B-A	DYR Ø 3.5 DYY Ø 4.5 DYB Ø 5.5 DYG Ø 7.0	PMA-DAN35/70-1	—	YES
MIS	C1 NP 3.3	S-02 (34685)		IO-S-02 (34689)	—	—	—	CN-MAN10	YES
	C1 SP 3.75, 4.2	S-03 (34686)		IO-S-03 (34690)	—	—	—	CS-MAN10	YES
	C1 WP 5.0	S-04 (34693)		IO-S-04 (34694)	—	—	—	CW-MAN10	YES
	M4 NP 3.3	B-04 (34678)		IO-B-04 (35273)	—	—	—	MN-MAN10	YES
	M4 SP 3.75, 4.2	B-01 (34527)		IO-B-01 (35212)	—	—	PMA-ZSV35-1	—	YES
	M4 WP 5.0, 6.0	B-02 (34528)		IO-B-02 (35213)	—	—	PMA-ZSV45-1	—	YES
	SEVEN NP 3.3	B-04 (34678)		IO-B-04 (35273)	—	—	—	MN-MAN10	YES
	SEVEN SP 3.75, 4.2	B-01 (34527)		IO-B-01 (35212)	—	—	PMA-ZSV35-1	—	YES
	SEVEN WP 5.0, 6.0	B-02 (34528)		IO-B-02 (35213)	—	—	PMA-ZSV45-1	—	YES
	V3 NP 3.3	S-01 (34684)		IO-S-01 (34688)	—	—	—	VN-MAN10	YES
	V3 SP 3.9, 4.3, 5.0	S-03 (34686)		IO-S-03 (34690)	—	—	—	CS-MAN10	YES
BioHorizons	Internal, Tapered Internal 3.5	B-01 (34527)	x	IO-B-01 (35212)	IO 5A-A	BIY	PMA-ZSV35-1	—	YES
	Internal, Tapered Internal 4.5	B-02 (34528)	x	IO-B-02 (35213)	IO 5A-B	BIG	PMA-ZSV45-1	—	YES
	Internal, Tapered Internal 5.7	B-03 (34529)	x	IO-B-03 (35214)	IO 5A-C	BIB	PMA-ZSV57-1	—	YES
	External Implant 3.5	A-03 (34518)		—	—	—	—	—	—
	External Implant 4.0	A-06 (34521)		—	—	—	—	—	—
	External Implant 5.0	A-09 (34524)		—	—	—	—	—	—

Please turn over 

All trademarks and company names are the property of their respective owners.

² The Atlantis IO FLO for Ankylos C/X is for single use only.

³ Use together with Elos lab screw.

⁴ Scanbody/Model analog exist but is not supported for Atlantis solutions.

Manufacturer	Implant	Model scanning		Intraoral scanning					
		Atlantis FLO		Atlantis IO FLO	Elos Accurate IO Scan Body	3Shape Scan Body Only Atlantis Implant Kit for 3Shape (DME)	Analog for Printed Models		Atlantis Printed Model
		Code (REF)	Kit Select	Code (REF)	REF	REF	Elos REF	MIS REF	Supported
Biomet 3i	Certain 3.25, 4/3	F-01 (34539)	x	IO-F-01 (35225)	IO 7B-A	BCP	PMA-BCE34-1	--	YES
	Certain 4.0, 5/4	F-02 (34540)	x	IO-F-02 (35226)	IO 7B-B	BCB	PMA-BCE41-1	--	YES
	Certain 5.0	F-02 (34540)	x	IO-F-02 (35226)	IO 7B-B	BCY	PMA-BCE50-1	--	YES
	Certain 6.0	F-02 (34540)	x	IO-F-02 (35226)	IO 7B-B	BCG	PMA-BCE60-1	--	YES
	Certain Prevail 3/4/3 (3.4), 4/3	F-01 (34539)	x	IO-F-01 (35225)	IO 7B-A	BCP	PMA-BCE34-1	--	YES
	Certain Prevail 4/5/4, 5/4	F-02 (34540)	x	IO-F-02 (35226)	IO 7B-B	BCB	PMA-BCE41-1	--	YES
	Certain Prevail 5/6/5, 6/5	F-02 (34540)	x	IO-F-02 (35226)	IO 7B-B	BCY	PMA-BCE50-1	--	YES
	Certain XP 4/5	F-02 (34540)	x	IO-F-02 (35226)	IO 7B-B	--	PMA-BCE50-1	--	YES
	Certain XP 5/6	F-02 (34540)	x	IO-F-02 (35226)	IO 7B-B	--	PMA-BCE60-1	--	YES
	Miniplant 3.25	A-07 (34522)		--	--	BEP	--	--	--
	Osseotite 3.75, 4.0	A-07 (34522)		--	--	--	--	--	--
	Osseotite 5.0, 6.0	A-07 (34522)		--	--	--	--	--	--
	Osseotite XP 4/5	A-07 (34522)		--	--	BEY	--	--	--
	Osseotite XP 5/6	A-07 (34522)		--	--	BEG	--	--	--
Camlog	CONOLOG 3.3	Q-01 (34666)		IO-Q-01 (35267)	--	--	--	--	--
	CONOLOG 3.8	Q-02 (34667)		IO-Q-02 (35268)	--	--	--	--	--
	CONOLOG 4.3	Q-02 (34667)		IO-Q-02 (35268)	--	--	--	--	--
	CONOLOG 5.0	Q-03 (34668)		IO-Q-03 (35269)	--	--	--	--	--
	Screw-line Implant 3.3	N-01 (34584)		IO-N-01 (35239)	IO 9A-A	CCG	PMA-CAS33-1	--	YES
	Screw-line/Root-line Implant 3.8	N-02 (34585)		IO-N-02 (35240)	IO 9A-B	CCY	PMA-CAS38-1	--	YES
	Screw-line/Root-line Implant 4.3	N-03 (34586)		IO-N-03 (35241)	IO 9A-C	CCP	PMA-CAS43-1	--	YES
	Screw-line/Root-line Implant 5.0	N-04 (34587)		IO-N-04 (35242)	IO 9A-D	CCB	PMA-CAS50-1	--	YES
	Screw-line/Root-line Implant 6.0	N-04 (34587)		IO-N-04 (35242)	IO 9A-D	CCB	PMA-CAS60-1	--	YES
Keystone Dental	PrimaConnex SD 3.3, Genesis 3.8	M-01 (34589)		--	--	--	--	--	--
	PrimaConnex SD 3.5	M-01 (34589)		--	--	--	--	--	--
	PrimaConnex RD 4.0, Genesis 4.5	M-02 (34590)		--	--	--	--	--	--
	PrimaConnex RD 4.1	M-02 (34590)		--	--	--	--	--	--
	PrimaConnex WD 5.0, Genesis 5.5, 6.5	M-03 (34591)		--	--	--	--	--	--

Please turn over 


All trademarks and company names are the property of their respective owners.

² The Atlantis IO FLO for Ankylos C/X is for single use only.

³ Use together with Elos lab screw.

⁴ Scanbody/Model analog exist but is not supported for Atlantis solutions.

Manufacturer	Implant	Model scanning		Intraoral scanning					
		Atlantis FLO		Atlantis IO FLO	Elos Accurate IO Scan Body	3Shape Scan Body Only Atlantis Implant Kit for 3Shape (DME)	Analog for Printed Models		Atlantis Printed Model
		Code (REF)	Kit Select	Code (REF)	REF	REF	Elos REF	MIS REF	Supported
Nobel Biocare	NobelActive 3.0	K-03 (34665)		IO-K-03 (35266)	IO 2B-D SA	NA3	PMA-NBA30-1	—	YES
	NobelActive / NobelReplace Conical Connection NP 3.5	K-01 (34578)		IO-K-01 (35233)	IO 2B-A SA	NAN	PMA-NBA35-1	—	YES
	NobelActive / NobelReplace Conical Connection RP 4.3	K-02 (34579)		IO-K-02 (35234)	IO 2B-B SA	NAR	PMA-NBA43-1	—	YES
	NobelActive / NobelReplace Conical Connection RP 5.0	K-02 (34579)		IO-K-02 (35234)	IO 2B-B SA	NAR	PMA-NBA43-1	—	YES
	NobelActive WP 5.5	K-04 (34677)		IO-K-04 (35272)	— ⁴	NAW	PMA-NBA60-1	—	YES
	NobelReplace NP 3.5 (Tri-channel connection)	D-01 (34533)	x	IO-D-01 (35219)	IO 2A-A	NRN	PMA-NBR35-1	—	YES
	NobelReplace RP 4.3 (Tri-channel connection)	D-02 (34534)	x	IO-D-02 (35220)	IO 2A-B	NRR	PMA-NBR43-1	—	YES
	NobelReplace WP 5.0 (Tri-channel connection)	D-03 (34535)	x	IO-D-03 (35221)	IO 2A-C	NRW	PMA-NBR50-1	—	YES
	NobelReplace 6.0 (Tri-channel connection)	D-04 (34536)	x	IO-D-04 (35222)	IO 2A-D	NR6	PMA-NBR60-1	—	YES
	Brånemark System NP 3.3	A-04 (34519)		—	IO 6A-A	NBN	PMA-BRA35-1	—	YES
	Brånemark System RP 3.75, 4.0	A-06 (34521)		—	IO 6A-B	NBR	PMA-BRA41-1	—	YES
	Brånemark System WP 5.0, 5.5	A-11 (34526)		—	IO 6A-C	NBW	PMA-BRA51-1	—	YES
Osstem	Osstem TS Mini 3.5	R-01 (34671)		IO-R-01 (35270)	—	OTM	—	—	—
	Osstem TS Regular 4.0, 4.5, 5.0, 6.0, 7.0	R-02 (34672)		IO-R-02 (35271)	—	OTR	—	—	—
Straumann	Bone Level 3.3 NC (Narrow CrossFit)	H-01 (34542)	x	IO-H-01 (35228)	IO 4A-A	SBN	PMA-SBO33-1	—	YES
	Bone Level 4.1 RC (Regular CrossFit)	H-02 (34543)	x	IO-H-02 (35229)	IO 4A-B	SBR	PMA-SBO41-1	—	YES
	Bone Level 4.8 RC (Regular CrossFit)	H-02 (34543)	x	IO-H-02 (35229)	IO 4A-B	SBR	PMA-SBO41-1	—	YES
	Standard/Standard Plus 4.8 RN (Regular Neck)	E-01 (34537)		—	IO 4B-A	SSR	PMA-SSY48-1	—	YES
	Standard/Standard Plus 4.8 WN (Wide Neck)	E-02 (34580)		—	IO 4B-B	SSW	PMA-SSY65-1	—	YES
	Standard Plus 3.3 NN (Narrow Neck)	G-01 (34541)		—	— ⁴	— ⁴	—	—	—
Zimmer Dental	Screw-Vent / Screw-Vent friction fit 3.3, 3.7	B-01 (34527)	x	IO-B-01 (35212)	IO 5A-A	—	PMA-ZSV35-1	—	YES
	Screw-Vent / Screw-Vent friction fit 4.5	B-02 (34528)	x	IO-B-02 (35213)	IO 5A-B	—	PMA-ZSV45-1	—	YES
	Tapered Screw-Vent / Tapered Screw-Vent friction fit 3.5	B-01 (34527)	x	IO-B-01 (35212)	IO 5A-A	ZTG	PMA-ZSV35-1	—	YES
	Tapered Screw-Vent / Tapered Screw-Vent friction fit 4.5	B-02 (34528)	x	IO-B-02 (35213)	IO 5A-B	ZTP	PMA-ZSV45-1	—	YES
	Tapered Screw-Vent / Tapered Screw-Vent friction fit 5.7	B-03 (34529)	x	IO-B-03 (35214)	IO 5A-C	ZTY	PMA-ZSV57-1	—	YES
	SwissPlus 4.8	E-01 (34537)		—	—	—	—	—	—

Please turn over 

All trademarks and company names are the property of their respective owners.

² The Atlantis IO FLO for Ankylos C/X is for single use only.

³ Use together with Elos lab screw.

⁴ Scanbody/Model analog exist but is not supported for Atlantis solutions.

Atlantis®

Atlantis® FLO Kit Select

Includes FLOs for the most frequently scanned interfaces, Atlantis FLO case and Atlantis FLO screwdriver (these may also be ordered separately, see below.) See above compatibility chart for further information on which FLOs are included in the kit.

Product	REF
Atlantis FLO Kit Select	34603
Atlantis FLO case	34602
Atlantis FLO screwdriver	34601
Atlantis FLO tag set	34664

